

PICARRO

SAM-C Installation and Startup Manual



Picarro Inc.
3105 Patrick Henry Drive
Santa Clara, CA 95054, USA
Phone: +1 408 962 3900 • Fax +1 408 962 3200
www.picarro.com

Document Number 40-0099 Revision A

This manual is the original instruction provided in English.



Picarro Notices

Picarro has prepared this manual solely for the information of its customers and as a general guide for the selection, installation, operation, and maintenance of the products described. Each application for these products is unique, and Picarro is willing to assist customers in their assessment and evaluation of Picarro products for specific uses and operating environments.

Please contact Picarro for questions regarding specific applications and additional information.

DISCLAIMER AND RESERVATION OF RIGHTS

Picarro has prepared this manual solely for the information and use by its customers as a guide for the selection, installation, operation, and maintenance of the products described.

EXCEPT AS PROVIDED IN THE TERMS AND CONDITIONS OF SALE FOR PICARRO PRODUCTS, PICARRO ASSUMES NO LIABILITY WHATSOEVER, AND PICARRO DISCLAIMS ANY EXPRESS OR IMPLIED WARRANTY, RELATING TO SALE OR USE OF PICARRO PRODUCTS, INCLUDING LIABILITY OR WARRANTIES RELATING TO MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.

Picarro reserves the right to change or update the contents of this manual and the specifications of its products at any time, without notice. Picarro has endeavored to include information that is current and accurate as of the date of the publication or revision of this document, but Picarro does not guarantee that this document is error free or that it is accurate with regard to any particular specification.

Picarro expressly reserves all intellectual property rights, including all intellectual property rights relating to any product described in this manual. This document does not grant any license, express or implied, by estoppel or otherwise, to any intellectual property rights of Picarro or any third party.

PATENTS

The products described in this manual are subject to Picarro patents and patents pending. Information about Picarro patents applicable to these products is available at www.picarro.com/company/patents.

TRADEMARKS

Picarro and the Picarro logo are trademarks of Picarro, Inc.

COPYRIGHT

Copyright © 2023 Picarro, Inc. All rights reserved.

Contact Information

Please contact Picarro for questions regarding specific applications and additional information.

General Technical Support:

Email: support@picarro.com

Phone: + 1 408 962 3991

European Technical Support:

Email: support@picarro.com

Phone: +31 85 888 1650

Customer Service:

Email: orders@picarro.com

Phone: + 1 408 962 3992

Table of Contents

Picarro Notices 2

Contact Information 3

Table of Contents 4

List of Figures 6

List of Tables 7

1. Introduction..... 8

 1.1 Intended Use 8

 1.2 Key Features 9

 1.3 How SAM-C Works..... 10

 1.4 Audience and Manual Scope 11

 1.5 Manual Purpose and Organization..... 11

 1.6 SEMI System Software 11

 1.7 Equipment Ratings/Specification Table 12

2. Safety Information, Warnings, and Hazards..... 14

 2.1 Advisory Notice – General Use 14

 2.2 Warning/Information Symbols and Text Conventions 14

 2.3 Warning Symbols on the System 15

 2.4 Warning Label Locations 16

 2.5 Seismic Protection Provisions..... 17

 2.6 Environmental Impact/Suitability for Recycling 18

3. Hardware Setup Prerequisites..... 20

 3.1 Personal Protective Equipment..... 20

 3.2 General Prerequisites 20

 3.3 Clean and Reference Gases..... 21

4. Hardware Equipment Installation 23

 4.1 Visual Inspection..... 23

 4.2 SAM-C Equipment – Key Components..... 24

 4.3 Analyzer Installation and Setup..... 27

 4.4 Electrical Connections 28

- 5. Hardware Equipment Operation 30**
 - 5.1 Power Up Sequence 30
 - 5.2 Power Down Sequence 30
 - 5.3 SAM-C Emergency Off Sequence 32
 - 5.4 SAM-C Emergency Stop Reset Sequence 32
- 6. SAM-C SEMI System Software User Manual 33**
- 7. Troubleshooting 34**
- 8. Equipment Maintenance and Service 36**
 - 8.1 Preventive Maintenance and Replacement of Consumable
Materials 36

List of Figures

Figure 1: SAM-C Integrated System – Front Panel Closed and Opened	8
Figure 2: System Diagram.....	11
Figure 3: Warning Label Locations.....	16
Figure 4: Anti-tipping Foot and Illustration of Overturning and Lateral Forces.....	17
Figure 5: Clean, Reference Gas, and Exhaust Ports at Rear Panel.....	22
Figure 6: SAM-C System – Front Panel Door Removed	23
Figure 7: SAM-C System Components (Front View)	24
Figure 8: SAM-C System Front Panel Components.....	24
Figure 9: SAM-C System Components (Back View)	25
Figure 10: SAM-C Right Side View	26
Figure 11: SAM-C Left Side View.....	27
Figure 12: Emergency Stop Switch Location.....	32

List of Tables

Table 1: Equipment Ratings and Specifications	12
Table 2: SAM-C System Analyzer Model Codes	13
Table 3: Warning/Information Icon Types.....	14
Table 4: Seismic Protection Calculations	18
Table 5: Summary of Material Construction Breakdown	19
Table 6: Problem Troubleshooting	35

1. Introduction

1.1 Intended Use

SAM stands for Sample, Analyze, Monitor.

SAM-C is a fully integrated Airborne Molecular Contamination (AMC) monitoring system with hardware and software working seamlessly together, enabling efficient and effective monitoring of AMCs in your fab.

AMCs can affect your final product and equipment. With contaminants coming from inside and outside the fab, it is vital to monitor and understand changes to the ambient environment. SAM-C monitors your fab conditions, using integrated sensors to track AMC trends in key fabrication areas; the system actively and accurately visualizes, analyzes, and stores data continuously, so you can view conditions and evaluate them according to your unique needs.



WARNING

The SAM-C and integrated analyzers are not intended to process hazardous materials at levels higher than 25% of their TLV (Threshold Limit Value) or LEL (Lower Explosive Limit). The system is only suitable for connection of non-HPM (Hazardous Production Material) calibration and reference gases.



Figure 1: SAM-C Integrated System – Front Panel Closed and Opened

1.2 Key Features

SAM-C offers the following elements as prime advantages in AMC detection:

- **FAST SYSTEM RESPONSE:** The gas-handling hardware of the SAM-C system has been specifically designed to minimize the “time-to-detection” response function of the sampling system. SAM-C’s flow control hardware enables you to monitor up to 32 unique locations throughout your fab and can rapidly detect concentration changes within the various monitored environments.
- **A USER-FRIENDLY GUI:** While the SAM-C hardware system is complex, its GUI is simple, easy-to-use, and configurable. From the SAM-C user interface, you can evaluate the concentration of one species or multiple species, and you can assess conditions of specific ports, multiple ports, specific instruments, multiple instruments, a bank or the entire fab. In addition, these conditions can be reproduced in both graph and tabulated forms, so you can see patterns and changes and generate reports for offline analysis and archiving.
- **MOBILE MANAGEMENT:** For network connected installations, SAM-C enables you to access all of its software features and have full system control from your desktop; you do not have to gown-up and enter the fab to modify measurement schedules and adjust measurement conditions. SAM-C also dynamically detects hardware connections and configures itself; you can easily validate these connections via the System page on the GUI.
- **EASY-TO-USE DATA EVALUATION TOOLS:** The SAM-C GUI enables you to track current conditions and to track the history of conditions. SAM-C is always monitoring and collecting concentration data on multiple species from each port in your fab. You can view a history of changes and spot trends, by species, by location, and by specified time period. This enables you to respond to AMC conditions in the moment and make improvements based on your analysis of historical trends.
- **SIMPLE RECIPE CREATION and SCHEDULING:** SAM-C provides **Schedule** and **Recipe Library** interfaces where you can easily schedule existing recipes (and view recipes that are currently running or scheduled), create and store new recipes, and edit existing recipes from the library.
 - **Recipe Creation Interface:** The user can create, run and store new recipes from either the Schedule or Library tabs. From the library tab, the user can also schedule, view, edit, or delete existing recipes.

The Recipe Creation interface enables you to create, customize, save, edit, and load recipes that automate monitoring of your fab quickly and easily. From the interface, you can add ports in sequential steps, run a reference gas through the system, and execute clean cycles during recipe runs. Each time a recipe is created, it is saved in the Recipe Library for easy access to edit as needed. It also keeps track of the last active recipe

and automatically recovers it in the event of a power-loss, resuming fab monitoring immediately.

- **Recipe Scheduling Interface:** The user can add recipes to the schedule to run at a future date and time, schedule as recurring, or run a recipe immediately. It is designed so the user does not need to create and run recipes manually every time.
- **Recipe Editing Interface:** The **View/Edit Recipe** interface enables you to open existing recipes to review and modify. This interface has the same functionality and look as the Create New Recipe interface but allows you to save the same recipe name as a different version, so the original recipe is preserved in the Recipe Library.

1.3 How SAM-C Works

Picarro's SAM-C maximizes the overall effectiveness of its AMC monitoring equipment.

First and foremost, Picarro's SAM-C leverages the excellent responsiveness of Picarro's gas analyzers. Picarro analyzers effectively and rapidly respond to any change in gas concentration, up or down. Most multiport sampling systems – placed upstream of gas analyzers – decrease the overall system responsiveness. This problem is compounded with each additional meter of upstream sample tubing. If these inherent issues are unaddressed by the sampling system design, events in your fab will go unnoticed and, potentially, you may respond to conditions that may be hours, if not days, in the past.

With SAM-C's specialized design, Picarro has made great strides in reducing, and in some cases eliminating these effects entirely. It enables you to monitor up to 32 sample lines with high throughput and use pressurized clean and reference gases to run periodic checks on analyzer baseline and calibrations. This enables you to respond immediately to changes in your fab's environment.

In addition, SAM-C's software supports the monitoring system with a simple and easy-to-use GUI. You can manually control the sampling hardware and make long-term test recipes that can be looped indefinitely. The software continuously collects and indexes all pertinent data and enables you to analyze it in any number of configurations. It collects and stores data by location, time frame, and species. You can also adjust recipes at any time from any place, with its portable and remote application option.




Figure 2: System Diagram

1.4 Audience and Manual Scope

This manual is designed for technicians, production managers, fab managers, and those involved in the monitoring and management of semiconductor fabrication facilities.

1.5 Manual Purpose and Organization

This manual begins with a brief introduction and then provides compliance and safety requirements and basic setup and start-up instructions. Most of the setup, installation and maintenance of SAM will be done by a manufacturer-trained service partner. Once the system has been installed and started, The user will refer to the included software operation manual.

1.6 SEMI System Software

Refer to the **SEMI Vx.x System Software User Manual** (Picarro PN 40-0096) included with this system for instructions on system operation.

1.7 Equipment Ratings/Specification Table

Table 1: Equipment Ratings and Specifications

Parameter	Value
Power Ratings	220-240 VAC 50/60Hz 3.1 kW Max.
Degrees of Ingress Protection	IPx0
Environmental Range for Equipment	
Indoor Use	Indoor Use Only
Altitude	10,000 Feet Operating
Operating Temperature	10 °C to 30 °C (operating) -10 °C to 50 °C (storage)
Maximum Relative Humidity	<85% R.H. non-condensing
Mains Supply Voltage Fluctuation	Up to ±10 % of the nominal voltage
Equipment Size – Standard (H x W x D)	2012 mm x 864 mm x 1172 mm (Depth is with keyboard opened) 79 in x 34 in x 46in
Maximum Weight	351.5 kg (775 lbs.)
Sampling Line	1/2-inch OD x 3/8-inch ID UHP-PFA tubing
Communication	Ethernet remote interface
Model Numbers	SAM-C-xx-yyyy “xx” indicates the number of ports (08, 16, 24, or 32) “yyyy” consists of alpha codes which indicate the analyzer models installed in your SAM system (See Table 2 below for model code matrix). For example: SAM-C-32-ABG0 is a 32 port system containing analyzer models SI3401, SI2104, and SI5450.

Table 2: SAM-C System Analyzer Model Codes

Code	Analyzer Model	Gases Measured
A	SI3401	NH ₃ , HF, HCl
B	SI2104	H ₂ S
C	SI2205	HF
D	SI2108	HCl
E	SI2306	HF, NH ₃
F	SI2103	NH ₃
G	SI5450	SO ₂

2. Safety Information, Warnings, and Hazards

2.1 Advisory Notice – General Use









Warning

Using SAM-C in a manner not specified by Picarro may result in damage to SAM and render it unsafe to operate.

2.2 Warning/Information Symbols and Text Conventions

Icon notes and warnings are used throughout this manual to provide information on dangers to either yourself or to the analyzer. The purpose of these icons is to provide a visual convention to alert you important information. They indicate dangers to either the operator or to the analyzer, and other important information.

Table 3: Warning/Information Icon Types







Icon	Description
 NOTE	NOTE indicates valuable information that you should be aware of before proceeding.
 WARNING	LASER WARNING alerts you of a laser danger.
 DANGER	DANGER indicates an imminently hazardous situation that, if not avoided, will result in death or severe injury.
 WARNING	WARNING indicates a potentially hazardous situation which, if not avoided, could result in death or severe injury.
 CAUTION	CAUTION alerts user of a potential danger to equipment or to the user.
 WARNING	HAZARDOUS VOLTAGE alerts user to areas that may expose a user to electrical energy that is high enough to cause injury or death.

Text Conventions

- *Italic* text identifies screen names and to emphasize important text, section reference links, or certain features.
- **Bold** text is for actions to take (such as clicking on or touching a UI button), cautions, statements, and text you should type or select in screens.

2.3 Warning Symbols on the System

The following symbols appear as labels on the SAM-C system as needed to indicate potential hazards.

Symbol	Meaning
	<p>HOT SURFACE: This indicates easily touched surfaces that can exceed normal temperature conditions.</p>
	<p>HAZARDOUS VOLTAGE: This shows an area on the machine that when touched, may cause an electric shock or a burn.</p>
	<p>ASPHYXIATION WARNING: This indicates an area where an oxygen deficient area exists and poses a danger of asphyxiation.</p>
	<p>LIVE AFTER EMO: This indicates an area where risk of electrical shock or burn is still present, even after an EMO switch has been pressed to disconnect power.</p>
	<p>ELECTRICAL LOCKOUT POINT: This indicates the location to place an electrical lockout device to prevent accidental electrical energizing of the SAM-C system.</p>
	<p>SEISMIC ANCHOR POINT: Indicates seismic anchor placement locations.</p>

2.4 Warning Label Locations




Figure 3: Warning Label Locations

2.5 Seismic Protection Provisions

To prevent the SAM-C unit from tipping over during a seismic event, each SAM-C is equipped with four anti-tipping feet. Each foot is constructed of welded 12-gauge cold rolled steel and bolted to the side of the unit with four 10 mm long stainless steel M6 hex head bolts. Each foot is anchored to the ground with two flanged stainless steel M12 hex head bolts.

The following illustrations and table define the measures that were taken to prevent the SAM-C from tipping over during a seismic event.

Based on the calculations of the overturning and lateral forces, each anchor bolt is subjected to a worst-case tensile force of 116.85 lbf and a lateral force of 59.1 lbf.




Figure 4: Anti-tipping Foot and Illustration of Overturning and Lateral Forces

Table 4: Seismic Protection Calculations

Variable	Notes	Value
W	Maximum normal operating weight of SAM-C	750 lbf
F_H	Maximum anticipated seismic horizontal force acting on SAM-C	0.63W
W_E	Effective weight of SAM-C accounting for vertical seismic force	0.74W
G	The center of gravity of SAM-C	As indicated
Z	Height of center of gravity	55 in
R	Tensile force on an anchor bolt	116.85 lbf
N	Total number of anchor bolts	8
n	Number of anchor bolts on one side of the SAM-C	4
X	Distance between two opposing sets of anchor bolts	35.83 in
L	Shortest distance from a set of anchor bolts to G	16.65 in

2.6 Environmental Impact/Suitability for Recycling

Hazardous Material Content

The European Union has identified certain substances that are considered environmentally hazardous according to the WEEE (Waste for Electrical and Electronic Equipment) and RoHS Directives. At the end of life of the instrument, the SAM-C system shall be dismantled, and the hazardous materials shall be identified, sorted, and collected.

The WEEE directive sets minimum standards for recycling of electrical and electronic waste. Those standards must be met. Local practice may exceed the minimum standards.

The crossed out wheeled bin symbol represents that hazardous content is included in the equipment. The equipment and parts with hazardous content shall not be disposed of with unsorted municipal waste. It is required that electrical and electronic equipment be disposed of under separate collection.



Hazardous Material in Electronic Equipment

- Lead is typically found in solder connections including PCBA terminations.
- Mercury can be found in electronic components such as switches, and relays.
- Lead, mercury, and cadmium may be found in insulation of electrical cables.
- Hexavalent Chromium may be used as a coating on frames, screw, and fasteners.
- All parts with hazardous material shall be separated and disposed of to comply with applicable local laws and regulations including with European WEEE Directive.

Summary of Material Construction Breakdown

The following table shows the overall approximate percentage by weight of material to show suitability of recycling, based on the SAM-C-32 configuration. The total weight of the system is approximately 570 pounds without the analyzers.

Table 5: Summary of Material Construction Breakdown

Material	Weight (lbs.)	Percentage
Steel	233.1	40.90%
Aluminum	207.7	36.40%
Stainless Steel	68.1	11.90%
PTFE	18.8	3.30%
Plastic (Nylon)	8.2	1.40%
Brass	7.8	1.30%
Zinc Alloy	7.2	1.20%
PBT	6.8	1.20%
PFA	5	0.90%
Copper	3	0.50%
FR-4 Fiber Epoxy Resin	1.3	0.20%
Clear Polycarbonate	1	0.20%
Neoprene	0.3	0.05%
Delrin	0.2	0.04%

3. Hardware Setup Prerequisites

3.1 Personal Protective Equipment

Any protective equipment shall be used in accordance with the instructions provided by the protective equipment supplier and as appropriate to the conditions in your fab.

3.2 General Prerequisites

- **Space Requirements:** The space requirements are 850 mm width, 2110 mm depth (with doors open), and 2020 mm height.
- **Space Requirements for Service:** The SAM-C needs to have at least 1m space on the right side of the chassis for servicing. If this is not practicable, servicing will require the system to be switched off, unanchored, and rolled to a more open location.
- **Power Requirements:** SAM-C is designed to operate with a 220-240 VAC power supply. If only a 110-120 VAC power supply is available, then a step-up transformer is required.

In the case of industrial field use, power the Picarro analyzer and sampling system from a clean and steady power source. Avoid intermittent electrical fluctuations or electrical power surges, as they can damage the Picarro analyzer and sampler electronics.

- **Step-up Transformer Requirements:** If required, Picarro recommends the following step-up transformer specification: 4000-Watt, 220-240 VAC, 35-Amp fuse.
- **Gas Cylinder Requirements:** Use chains and/or a cage to securely hold compressed gas cylinders in place. Be sure the setup meets safety requirements as outlined by the gas supplier as well as the health and safety codes at the installation site.
- **PFA Tubing:** Use 1/2-inch outer diameter and 3/8-inch inner diameter PFA tubes, as much as required for the sample lines. Use 3/8-inch outer diameter PFA tubes for Reference and Clean ports.
- **Ventilation:** The customer is responsible for safely venting potentially toxic exhaust from all pumps.

3.3 Clean and Reference Gases

This section applies only to users who want to supply external Clean and Reference gases to the SAM-C system.

Gas Supply Requirements

- Regulated to 20-25 PSIG
- Flow capacity greater than 50 SLPM
- Gas input ports are 1/2-inch OD push-to-connect fittings. Recommended tubing is 1/2-inch OD x 3/8-inch ID, PFA tubing.

Recommended Clean Gases

- House Nitrogen, N₂
- Clean Dry Air (CDA)



CAUTION

When nitrogen or any other gases that are not 100% air are connected to the system, the SAM-C needs to be installed in a well-ventilated location.

Doors and panels need to be opened for ventilation prior to entering SAM-C.

Viable Surrogate Reference Gases

- Use a CO₂ Bottle ~1% to validate NH₃ measurement.
- Use a CH₄ Bottle ~100 ppm to validate HCl measurement.
- Use Zero Air cylinder ~21% O₂ to validate HF measurement.



NOTE

The line from the Clean source must be plugged into the “CLEAN” fitting, and the line from the Reference Gas tank must be plugged in to the “REF” fitting. See Figure 5 below.




Figure 5: Clean, Reference Gas, and Exhaust Ports at Rear Panel

4. Hardware Equipment Installation

While it is likely that the SAM-C hardware will be installed by a manufacturer trained service partner, the following guidelines are Picarro's recommendations for installation. Be aware of your fab safety requirements during installation.



CAUTION

The safety of your overall system is the responsibility of your onsite staff.



NOTE

Short circuit current rating (SCCR) of the SAM-C is 5 kA.

4.1 Visual Inspection

A manufacturer trained service partner will install the system and perform the following visual inspection.



Prior to powering on the system, inspect the following items:

1. The sampling system's physical condition and the overall condition of the system
2. The sampling ports and pneumatic connections
3. The analyzer pumps (240V) and cooling fan
4. The emergency electric safety switch, the integrity of the wiring connections, and the switches

Figure 6: SAM-C System – Front Panel Door Removed

4.2 SAM-C Equipment – Key Components




Figure 7: SAM-C System Components (Front View)




Figure 8: SAM-C System Front Panel Components




Figure 9: SAM-C System Components (Back View)



Figure 10: SAM-C Right Side View



NOTE

The front and back panels are hinged for easy opening. However, the side panels are not hinged and must be removed for access. A manufacturer-trained service partner should only remove them.




Figure 11: SAM-C Left Side View

4.3 Analyzer Installation and Setup

1. Install the front rack-mount onto the analyzer.
2. Make the following connections to the analyzer (back side) area.



3. Connect the power cord to each analyzer and ensure that the power switch is in ON position.



Note: The Power plug locks when inserted. To remove, slide the red release tab back and pull.



4. Connect the sample line to the sample port.



5. Connect the external pump line to the exhaust port.



6. Connect the HDMI cable plus DVI adapter (along with USB) to the DVI port.



7. Do NOT connect the USB cable to USB-3 ports on the analyzer.



8. Connect the RJ45 ethernet cable in the network port (labeled) into one of the network sharing ports on the back side of the analyzer. Either port will work.

4.4 Electrical Connections



1. Ensure that all the power cables are seated properly. Check the sampling controller back panel, all analyzer back panels, and the external power distribution unit for the power cord connection.



2. Ensure that the EMERGENCY OFF button is in the release position.



3. Double-check that each Analyzer pump voltage is set to 240-VAC. All the Analyzer pumps are shipped with the 240-VAC setting as default.



4. Connect the sampling system (220-240 VAC EU type connector) to the 220-240 VAC power supply outlet (rated 20A). If connecting with a step-up transformer, connect the transformer to a 110-120 VAC power supply (rated 35A) and connect the sampling system electrical connection to the step-up transformer 220-240 VAC outlet.



Note: The Inlet plug is supplied with SAM-C but the connector on the user end is not.



Caution: Ensure that the step-up transformer provides a ground connection to the SAM-C. Picarro recommends consulting an electrician if there is any doubt on how to connect the transformer for safe operation."

5. Hardware Equipment Operation

5.1 Power Up Sequence

1. Uncoil the AC power supply cord and plug the connector into an appropriate 220-240 VAC wall socket (or step up transformer in areas with 110-120 VAC supply).



2. Press the ON switch on the front face of the SAM-C chassis. The power up sequence is automatic once the ON switch is pressed.

5.2 Power Down Sequence

Power Down sequence requires shutting down items in the following order.


- Analyzer 1, Analyzer 2, Analyzer 3, Analyzer 4 (as applicable)
- The SAM-C Computer
- The SAM-C Main Power Supply

The analyzers and SAM-C Computer can be shut down either from the User interface (the primary method) or from the keyboard and KVM switch using hot keys (the alternative method).

Powering Down Analyzers and SAM-C PC via the User Interface



1. At the bottom of the Navigation Panel, click the **Power** button.
2. A pop-up window displays. Click the **Shutdown** button.
3. This shuts down each analyzer one at a time, then shuts down the SAM-C computer. Once this is done, to power the system back up, you will need to follow section **5.1, Power Up Sequence**.



NOTE

Refer to the SEMI Vx.x System Software User Manual (Picarro PN 40-0096) for full system operation instructions.

Powering Down Analyzers and SAM-C PC via KVM Switch and Keyboard


If the need arises to shut the analyzers and SAM-C PC down without using the User Interface, use the following steps.

Analyzer 1 - 4 – KVM Port Power Down Sequence

A KVM switch is used to toggle between the SAM-C and internal analyzers. You will use the keyboard hot keys to Power Down.



(Press twice)





1. Press the hot keys on the key board. Press [Ctrl], [Ctrl] (Control button is pressed twice in succession), [#], [ENTER] where [#] is [2] - [5] for the four analyzers. As you proceed, each individual analyzer GUI will appear, from which you can shut each one analyzer.
2. Follow the instructions in the Analyzer user manual to properly shut down each analyzer.
3. Repeat these steps for any additional analyzers, changing the KVM port number pressed based on KVM position.



NOTE

If the hot keys are not functioning, toggle the KVM Switch manually. The Switch can be found by opening the door at the front of the SAM-C. Use the buttons on the front of the KVM.




SAM-C Computer Power Down Sequence

4. Use the hot keys to switch to the SAM-C PC. Using the keyboard press [Ctrl], [Ctrl] (Control button is pressed twice in succession), [1], [ENTER]. You will see the SAM-C computer GUI.



5. Press the **cmd** key on your keyboard. This brings up the SAM-C computer desktop.



6. On the desktop, scroll up to the Power button (upper-right) and press the **Power** button from the drop-down list.

7. A pop-up window displays, click the **Power Off** button.

SAM-C Main Supply Power Down



Once the SAM-C computer is shut down, press the red **OFF** button on the front of SAM-C. This turns off AC power to the rest of the system.

5.3 SAM-C Emergency Off Sequence



The EMERGENCY OFF switch is located on the front left side of the SAM-C (Figure 12).

1. Press the switch to activate the emergency shut down sequence.
2. Once activated, all power will be cut immediately.




Figure 12: Emergency Stop Switch Location

5.4 SAM-C Emergency Stop Reset Sequence



1. To reset the EMERGENCY OFF switch, twist the knob clockwise (in the direction indicated) to release.
2. Check that SAM-C is properly connected to a suitable VAC outlet (see the SAM-C Power Up Sequence).



3. Press the green ON button on the front of the SAM-C chassis. The ON indicator should light up in green.

6. SAM-C SEMI System Software User Manual

For user interface operation instructions, refer to the manual, **SEMI Vx.x System Software User Manual** (Picarro PN 40-0096) included in your SAM-C shipment.

7. Troubleshooting

Prior to beginning any troubleshooting, first read the following notes and warnings. After that, refer to **Table 6: Problem Troubleshooting** below for a listing of typical problems that may arise. Each problem provides inspection areas and indicators that may give clues to what the cause is and the corrective action to take.



WARNING

All troubleshooting that requires opening system panels shall be done by a manufacturer-trained service partner.



WARNING

Pressurized clean and reference gases may leak. Any isolation valves at the supply must be shut off, and doors and panels need to be opened for ventilation prior to entering.



CAUTION

Lockout of electrical power and/or compressed gas supplies need to be verified using multimeter/pressure gauge (as applicable).



CAUTION

The system must be shut off or individual modules switched off and unplugged prior to opening any panels for service.



WARNING

Regulatory Employer and Facility Procedures (such as NFPA 70E), must be followed by any personnel conducting troubleshooting in energized areas. Only trained and authorized personnel are allowed to perform troubleshooting in energized areas.



CAUTION

Noise levels between 70 dBA and 75 dBA have been measured when the panels are removed. Simple troubleshooting should take no more than 30 minutes.



WARNING

After the EMO switch is engaged and the system power is cut, the following components remain LIVE: circuit breaker, surge protector, line filter, 24VDC, 2.2A power supply.

Table 6: Problem Troubleshooting

Problem/Issue	Inspection/Indications	Corrective Action
Valve not switching between lines	Valve	Check the power connection and USB connection.
	Indicator light	
Slow response	Gas Leak	Check that the sample lines are connected properly.
	Tube blockage	Check the sample line for any blockage.
	Pump operation degraded	Check to see if the sample pump is functioning by feeling for exhaust.
	Operation sequence	Check that the operation sequence has been defined adequately and is running.
No power	Electrical cable connection	<p>Check the breaker at the back of the SAM. The circuit breaker can be probed through the cover with 1.5 mm diameter test probe tips.</p> <p>Check that the EMERGENCY OFF switch is not engaged.</p> <p>Check the integrity of the power connection on all sampling boxes.</p>
Keyboard not functioning	Keyboard USB connection	Ensure that only the keyboard USB is connected to the KVM unit and that all other USBs are connected to the rack integrated PC.
Mouse not functioning	Mouse USB connection	Ensure that only the mouse USB is connected to the KVM unit and that all other USBs are connected to the rack integrated PC.

8. Equipment Maintenance and Service

All service and maintenance shall be done by a manufacturer trained service partner. There are no customer serviceable parts on this system.

Please contact your manufacturer trained service partner to schedule service calls, and trained personnel will perform any required service or maintenance.

8.1 Preventive Maintenance and Replacement of Consumable Materials

Fan Assemblies and Pumps

Fans: Picarro recommends that cooling fan assemblies be replaced every 2 years.

Pumps: SAM-C pumps cannot be rebuilt on-site. Picarro recommends pump replacement every 2 years to avoid system downtime.